ADINSTRUMENTS

MLA1093 Abrasive Gel

Accessories

The MLA1093 Abrasive Gel 4oz tube can be used to prepare an EEG, EOG or EMG recording site. The mild abrasive in the gel helps lower impedances at the electrode site with minimal discomfort and irritation. Apply gel to electrode site with gauze pad or cotton swab..

Caution

Read “Statement of Intended Use” on our website or in “Getting Started with PowerLab” before use.

Ordering Information:

MLA1093 Abrasive Gel

A pack of 3 tubes of Abrasive Gel is also available: MLA1093B

WARRANTY: ADInstruments PowerLab data acquisition units (Product Number Prefix: PL), Front-end Signal Conditioners (Product Number Prefix: FE) are warranted against defects in materials and workmanship for a period of 5 years from the date of purchase. Other PowerLab data acquisition units and ADInstruments manufactured Front-end and Pod Signal Conditioners, and Instruments are warranted of a period of 3 years from the date of purchase. Third party products are covered by the manufacturer’s warranty. Warranties are void if the product has been damaged due to negligence. Consumables and electrodes are not covered by a warranty. All questions regarding service and warranty should be directed to your nearest ADInstruments representative.